




# लालबन्दी नगरपालिका नगर कार्यपालिकाको कार्यालय

लालबन्दी, सर्लाही  
मधेश प्रदेश, नेपाल



फोन नं. ०४६५०९१००  
०४६५०९३३३  
०४६५०९४४४

Email: mailalbandi@gmail.com

## दररेट उपलब्ध गराईदिने बारे सूचना

सूचना प्रकाशन मिति:- २०८०/१०/१२

यस कार्यालय मातहतका स्वास्थ्य संस्थाहरुको लागि तपसिल बमोजिमका उपकरणहरु आवश्यक भएकोले सार्वजनिक खरिद नियमावली, २०६४ नियम ११ बमोजिम संलग्न स्पेसिफिकेशन बमोजिमका उपकरणहरुको दररेट सम्बन्धित फर्म/कम्पनी/सप्लायर्स लाई ७ दिन भित्र यस कार्यालयमा उपलब्ध गराईदिनुहुन अनुरोध गरिन्छ।

### तपसिल


क्र. स.	उपकरणहरुको नाम	स्पेसिफिकेशन	ब्रान्ड/कम्पनि
१	Adult Weighing Scale	Super slim Electronic Glass Personal Adult Weighing Scale	Branded Company ISO 9001 or ISO 13485
२	Sphygmomanometer	Plam Type Aneroid Sphygmomanometer	Branded Company ISO 9001 or ISO 13485
३	Weighing salter types	Baby weighing scale salter types ABS Plastic body with Trouser	Branded Company ISO 9001 or ISO 13485
४	Stethoscope	Dual Head Stethoscope	Branded Company ISO 9001 or ISO 13485
५	Thermometer, Clinical, Digital	Thermometer, Clinical, Digital	Branded Company ISO 9001 or ISO 13485

प्रमुख प्रशासकीय अधिकृत  
२०८०



TSN:116131

## Adult Weighing Scale



Date:01.2024


S.N.	Purchaser's Specifications	
	<b>Super slim Electronic Glass Personal Adult Weighing Scale</b>	
	<b>Manufacturer</b>	
	<b>Brand</b>	Branded Company ISO 9001 or ISO 13485
	<b>Type / Model</b>	
	<b>Country of Origin</b>	
<b>1</b>	<b>Description of Function</b>	
1.1	An electrically operated adult weighing scale	
<b>2</b>	<b>Operational Requirements</b>	
2.1	Electronic Adult Weighing Scale (Platform Type)	
<b>3</b>	<b>System Configuration</b>	
3.1	Electronic Adult Weighing Scale (Platform Type) on main power as well as battery operated.	
<b>4</b>	<b>Technical Specifications</b>	
4.1	Capacity : 150 kg-180Kg	
4.2	Accuracy : 100 g	
4.4	The scale must be made up of heavy duty sturdy platform with powder coated frames.	
4.5	Display: LED / LCD : Big LED Size	
4.6	Auto-on Function	
4.7	HOLD function to lock the weight.	
4.8	Memory function, to keep the last weight in memory.	
4.9	The Scale must have inbuilt rechargeable battery backup for minimum of 8 hrs. with AC adapters.	
<b>5</b>	<b>Accessories, spares and consumables</b>	
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).	
<b>6</b>	<b>Operating Environment</b>	
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
<b>7</b>	<b>Standards and Safety Requirements</b>	
7.1	Must submit ISO 9001 or ISO 13485	
<b>9</b>	<b>Warranty</b>	
9.1	Comprehensive warranty for 2 year.	
<b>10</b>	<b>Maintenance Service During Warranty Period</b>	
10.1	During warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
<b>11</b>	<b>Installation and Commissioning</b>	
11.1	Supplier must accomplish proper commissioning of the equipment on site.	
<b>12</b>	<b>Documentation</b>	
12.1	User (Operating) manual in English	
12.2	Service (Technical / Maintenance) manual in English	
12.3	List of important spare parts and accessories with their part numbers and costing.	
12.4	Certificate of calibration and inspection from factory.	



TSN:116135

Approved Date:01.2024

## Plam Type Aneroid Sphygmomanometer





S.N.	Purchaser's Specifications	
	<b>Plam Type Aneroid Sphygmomanometer</b>	
	<b>Manufacturer</b>	
	<b>Brand</b>	Branded Company ISO 9001 or ISO 13485
	<b>Type / Model</b>	
	<b>Country of Origin</b>	
<b>1</b>	<b>Description of Function</b>	
1.1	Sphygmomanometer is a device used to measure blood pressure, composed of an inflatable cuff to restrict blood flow, and a mechanical manometer to measure the pressure. Super performance And Quality	
<b>2</b>	<b>Operational Requirements</b>	
2.1	Aneroid sphygmomanometer having a dial to show clear numbers and pointer / needle for measurement of pressure.	
<b>3</b>	<b>System Configuration</b>	
3.1	<ul style="list-style-type: none"> <li>• Aneroid sphygmomanometer</li> <li>• Cuffs for adult size (regular)</li> <li>• Inflation bulb</li> <li>• Carrying pouch</li> </ul>	
<b>4</b>	<b>Technical Specifications</b>	
4.1	Packed in easy carrying high quality pouch made of waterproof cloth to accommodate cuff, and inflation bulb.	
4.2	Gauge to be calibrated in 2 mm Hg units.	
4.3	Must provide blood pressure cuffs for child size and for adult size (regular).	
<b>5</b>	<b>Accessories, spares and consumables</b>	
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.	
<b>6</b>	<b>Operating Environment</b>	
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
<b>7</b>	<b>Standards and Safety Requirements</b>	
7.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007.	
<b>8</b>	<b>User Training</b>	
8.1	Not applicable.	
<b>9</b>	<b>Warranty</b>	
9.1	Warranty for 2year.	
<b>10</b>	<b>Maintenance Service During Warranty Period</b>	
10.1	Standard warranty conditions are applicable.	
<b>11</b>	<b>Installation and Commissioning</b>	
11.1	Must supply preassembled unit, ready to use.	
<b>12</b>	<b>Documentation</b>	
12.1	User's manual in English	



## Stethoscope

S.N.	Purchaser's Specifications	
	<b>Dual Head Stethoscope</b>	
	<b>Manufacturer</b>	
	<b>Brand</b>	Branded Company ISO 9001 or ISO 13485
	<b>Type / Model</b>	
	<b>Country of Origin</b>	
<b>1</b>	<b>Description of Function</b>	
1.1	The stethoscope is used for listening to the beating heart of a human, or the lungs. It is also used for listening to the flow of the blood in the surrounding area of the heart.	
<b>2</b>	<b>Operational Requirements</b>	
2.1	Dual type stethoscope - Physician's stethoscope.	
<b>3</b>	<b>System Configuration</b>	
3.1	<ul style="list-style-type: none"> <li>• Stethoscope, dual cup/bell</li> <li>• Tubes</li> </ul>	
<b>4</b>	<b>Technical Specifications</b>	
4.1	Dual, cup/bell and diaphragm head	
4.2	Head and ear tube assembly to be made of non-ferrous metal,	
4.3	Tubes to be synthetic material and ear tubes to have shaped plastic cushion ends.	
<b>5</b>	<b>Accessories, spares and consumables</b>	
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.	
<b>6</b>	<b>Operating Environment</b>	
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
<b>7</b>	<b>Standards and Safety Requirements</b>	
7.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007.	
<b>8</b>	<b>User Training</b>	
8.1	Not applicable.	
<b>9</b>	<b>Warranty</b>	
9.1	Warranty for 2year.	
<b>10</b>	<b>Maintenance Service During Warranty Period</b>	
10.1	Standard warranty conditions are applicable.	
<b>11</b>	<b>Installation and Commissioning</b>	
11.1	Must supply preassembled unit, ready to use.	
<b>12</b>	<b>Documentation</b>	
12.1	User's manual in English	



TSN:116132

Approved Date:01.2024



## Baby Scale, Weighing salter types


S.N.	Purchaser's Specifications
	<b>Baby weighing scale salter types ABS Plastic body with Trouser</b>
	<b>Manufacturer</b>
	<b>Brand</b> Indo surgical
	<b>Type / Model</b> Baby suspension Scale
	<b>Country of Origin</b>
<b>1</b>	<b>Description of Function</b>
1.1	Mechanical Infant and child Weighing Scale
<b>2</b>	<b>Operational Requirements</b>
2.1	baby weighing scale with two metal <b>Suspension, trouser or hanging types as well as electronic or battery operated scales are not acceptable.</b>
<b>3</b>	<b>System Configuration</b>
3.1	weighing scale <b>salter types 25kg</b>
<b>4</b>	<b>Technical Specifications</b>
4.1	Large pan, approx. 450mm long and 300mm wide.
4.2	Pan to have flat base with lips on sides only (NOT on ends).
4.3	Lips to be approx. 80mm high
4.3	Pan made of Acrylic, Moulded Engineering Plastic or Metal with an easy to clean surface
4.4	Scale to weigh 0 to 6 Kg in increments of 50g
4.5	Dial type or Danish yard-arm balance types accepted
4.6	Shall have Tare/Zero adjustment system
<b>5</b>	<b>Accessories, spares and consumables</b>
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.
<b>6</b>	<b>Operating Environment</b>
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.
<b>7</b>	<b>Standards and Safety Requirements</b>
7.1	Must submit ISO 9001 or ISO 13485:2003/AC:2007
<b>8</b>	<b>User Training</b>
8.1	Not applicable.
<b>9</b>	<b>Warranty</b>
9.1	Warranty for 1 year.
<b>10</b>	<b>Maintenance Service During Warranty Period</b>
10.1	Standard warranty conditions are applicable.
<b>11</b>	<b>Installation and Commissioning</b>
11.1	Must supply preassembled unit, ready to use.
<b>12</b>	<b>Documentation</b>
12.1	User (Operating) and Service (Technical/Maintenance) manuals to be supplied in English.



TSN:116141

## Thermometer, Clinical, Digital

Approved Date: 01/2024



Purchaser's Specifications	
S.N.	
	<b>Digital Thermometer, Clinical</b>
	<b>Manufacturer</b>
	<b>Brand</b>
	<b>Type/Model</b>
	<b>Country of Origin</b>
1	<b>Description of Function</b>
1.1	Clinical digital thermometer to check the body temperature.
2	<b>Operational Requirements</b>
2.1	Human oral use.
3	<b>System Configuration</b>
3.1	Digital Thermometer, Clinical.
4	<b>Technical Specifications</b>
4.1	Flat type, wide thermometer, safe to use, no glass, no mercury.
4.2	Scale: Celsius scale.
4.3	Measurement range: 32°C to 45°C
4.4	Accuracy: +/- 0.1°C between 35°C to 42°C.
4.5	Display: Liquid crystal display, easy to read.
4.6	Shall work on a battery.
4.7	Shall have a low battery indicator.
4.8	Shall have facility of beep sound and switch off.
4.9	Water proof for ease of cleaning.
5	<b>Accessories, Spares and Consumables</b>
5.1	Shall be provided with battery. Bidder to indicate the type of and number of battery to be supplied.
5.2	<b>Packing:</b>
	• Single piece packing in plastic barrel with cover.
6	<b>Operating Environment</b>
6.1	The system offered shall be designed to store and to operate normally under the conditions of the purchaser's country. The conditions include, Climate, Temperature, Humidity, etc.
7	<b>Standards and Safety Requirements</b>
7.1	CE (93/42 EEC Directives)
8	<b>User Training</b>
8.1	Not applicable.
9	<b>Warranty</b>
9.1	Warranty for 1 year after acceptance.
10	<b>Maintenance Service During Warranty Period</b>
10.1	Not applicable.
11	<b>Installation and Commissioning</b>
11.1	Must supply preassembled unit, ready to use.
12	<b>Documentation</b>
12.1	User's manual to be supplied in English.

*Brick*

*[Signature]*