




लालबन्दी नगरपालिका नगर कार्यपालिकाको कार्यालय

लालबन्दी, सर्लाही
मधेश प्रदेश, नेपाल



दरभाउपत्र पेश गर्ने बारे सूचना ।

सूचना प्रकाशन मिति:- २०८०/१०/२६

सार्वजनिक खरिद नियमावली, २०६४ नियम ८५ उपनियम ४ बमोजिम संलग्न स्पेसिफिकेशन बमोजिमका तपसिलका उपकरणहरूको दरभाउपत्र ७ दिन भित्र पेश गर्न जानकारी गरिन्छ । साथै दरभाउपत्र पेश गर्दा संस्था वा फर्म दर्ता प्रमाणपत्र, स्थायी लेखा नम्बर वा मूल्य अभिवृद्धि कर दर्ताको प्रमाणपत्र, कर चुक्ता प्रमाणपत्र, व्यवसायिक इजाजतपत्र र स्पेसिफिकेशन बमोजिमको उपकरणहरूको ISO 9001 प्रमाणपत्र र उपकरणको क्याटलग सहित खाम बन्द गरी लाहा छाप लगाएर अनिवार्यरूपमा पेश गर्नुहुन । ISO 9001 प्रमाणपत्र नभएको उपकरणहरू मूल्यांकन गरिने छैन ।

तपसिल


क्र.स.	उपकरणहरूको नाम	स्पेसिफिकेशन	
१	Adult Weighing Scale	Super slim Electronic Glass Personal Adult Weighing Scale	Branded Company ISO 9001 or ISO 13485
२	Sphygmomanometer	Plam Type Aneroid Sphygmomanometer	Branded Company ISO 9001 or ISO 13485
३	Weighing salter types	Baby weighing scale salter types ABS Plastic body with Trouser	Branded Company ISO 9001 or ISO 13485
४	Stethoscope	Dual Head Stethoscope	Branded Company ISO 9001 or ISO 13485
५	Thermometer, Clinical, Digital	Thermometer, Clinical, Digital	Branded Company ISO 9001 or ISO 13485

प्रमुख प्रशासकीय अधिकृत

०८०-५०

TSN:116131

Adult Weighing Scale




Date:01.2024

S.N.	Purchaser's Specifications	
	Super slim Electronic Glass Personal Adult Weighing Scale	
	Manufacturer	
	Brand	Branded Company ISO 9001 or ISO 13485
	Type / Model	
	Country of Origin	
1	Description of Function	
1.1	An electrically operated adult weighing scale	
2	Operational Requirements	
2.1	Electronic Adult Weighing Scale (Platform Type)	
3	System Configuration	
3.1	Electronic Adult Weighing Scale (Platform Type) on main power as well as battery operated.	
4	Technical Specifications	
4.1	Capacity : 150 kg-180Kg	
4.2	Accuracy : 100 g	
4.4	The scale must be made up of heavy duty sturdy platform with powder coated frames.	
4.5	Display: LED / LCD : Big LED Size	
4.6	Auto-on Function	
4.7	HOLD function to lock the weight.	
4.8	Memory function, to keep the last weight in memory.	
4.9	The Scale must have inbuilt rechargeable battery backup for minimum of 8 hrs. with AC adapters.	
5	Accessories, spares and consumables	
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).	
6	Operating Environment	
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
7	Standards and Safety Requirements	
7.1	Must submit ISO 9001 or ISO 13485	
9	Warranty	
9.1	Comprehensive warranty for 2 year.	
10	Maintenance Service During Warranty Period	
10.1	During warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
11	Installation and Commissioning	
11.1	Supplier must accomplish proper commissioning of the equipment on site.	
12	Documentation	
12.1	User (Operating) manual in English	
12.2	Service (Technical / Maintenance) manual in English	
12.3	List of important spare parts and accessories with their part numbers and costing.	
12.4	Certificate of calibration and inspection from factory.	

TSN:116135

Approved Date:01.2024



Plam Type Aneroid Sphygmomanometer



S.N.	Purchaser's Specifications	
	Plam Type Aneroid Sphygmomanometer	
	Manufacturer	
	Brand	Branded Company ISO 9001 or ISO 13485
	Type / Model	
	Country of Origin	
1	Description of Function	
1.1	Sphygmomanometer is a device used to measure blood pressure, composed of an inflatable cuff to restrict blood flow, and a mechanical manometer to measure the pressure. Super performance And Quality	
2	Operational Requirements	
2.1	Aneroid sphygmomanometer having a dial to show clear numbers and pointer / needle for measurement of pressure.	
3	System Configuration	
3.1	• Aneroid sphygmomanometer • Cuffs for adult size (regular) • Inflation bulb • Carrying pouch	
4	Technical Specifications	
4.1	Packed in easy carrying high quality pouch made of waterproof cloth to accommodate cuff, and inflation bulb.	
4.2	Gauge to be calibrated in 2 mm Hg units.	
4.3	Must provide blood pressure cuffs for child size and for adult size (regular).	
5	Accessories, spares and consumables	
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.	
6	Operating Environment	
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
7	Standards and Safety Requirements	
7.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007.	
8	User Training	
8.1	Not applicable.	
9	Warranty	
9.1	Warranty for 2year.	
10	Maintenance Service During Warranty Period	
10.1	Standard warranty conditions are applicable.	
11	Installation and Commissioning	
11.1	Must supply preassembled unit, ready to use.	
12	Documentation	
12.1	User's manual in English	

Stethoscope

S.N.	Purchaser's Specifications	
	Dual Head Stethoscope	
	Manufacturer	
	Brand	Branded Company ISO 9001 or ISO 13485
	Type / Model	
	Country of Origin	
1	Description of Function	
1.1	The stethoscope is used for listening to the beating heart of a human, or the lungs. It is also used for listening to the flow of the blood in the surrounding area of the heart.	
2	Operational Requirements	
2.1	Dual type stethoscope - Physician's stethoscope.	
3	System Configuration	
3.1	• Stethoscope, dual cup/bell • Tubes	
4	Technical Specifications	
4.1	Dual, cup/bell and diaphragm head	
4.2	Head and ear tube assembly to be made of non-ferrous metal,	
4.3	Tubes to be synthetic material and ear tubes to have shaped plastic cushion ends.	
5	Accessories, spares and consumables	
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.	
6	Operating Environment	
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
7	Standards and Safety Requirements	
7.1	Must submit ISO 9001 or ISO 13485:2003/AC: 2007.	
8	User Training	
8.1	Not applicable.	
9	Warranty	
9.1	Warranty for 2year.	
10	Maintenance Service During Warranty Period	
10.1	Standard warranty conditions are applicable.	
11	Installation and Commissioning	
11.1	Must supply preassembled unit, ready to use.	
12	Documentation	
12.1	User's manual in English	

TSN:116132

Approved Date:01.2024




Baby Scale, Weighing salter types

S.N.	Purchaser's Specifications
	Baby weighing scale salter types ABS Plastic body with Trouser
	Manufacturer
	Brand Indo surgical
	Type / Model Baby suspension Scale
	Country of Origin
1	Description of Function
1.1	Mechanical Infant and child Weighing Scale
2	Operational Requirements
2.1	baby weighing scale with two metal Suspension, trouser or hanging types as well as electronic or battery operated scales are not acceptable.
3	System Configuration
3.1	weighing scale salter types 25kg
4	Technical Specifications
4.1	Large pan, approx. 450mm long and 300mm wide.
4.2	Pan to have flat base with lips on sides only (NOT on ends).
4.3	Lips to be approx. 80mm high
4.3	Pan made of Acrylic, Moulded Engineering Plastic or Metal with an easy to clean surface
4.4	Scale to weigh 0 to 6 Kg in increments of 50g
4.5	Dial type or Danish yard-arm balance types accepted
4.6	Shall have Tare/Zero adjustment system
5	Accessories, spares and consumables
5.1	All standard accessories/consumables/parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.
6	Operating Environment
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.
7	Standards and Safety Requirements
7.1	Must submit ISO 9001 or ISO 13485:2003/AC:2007
8	User Training
8.1	Not applicable.
9	Warranty
9.1	Warranty for 1 year.
10	Maintenance Service During Warranty Period
10.1	Standard warranty conditions are applicable.
11	Installation and Commissioning
11.1	Must supply preassembled unit, ready to use.
12	Documentation
12.1	User (Operating) and Service (Technical/Maintenance) manuals to be supplied in English.

TSN:116141

Thermometer, Clinical, Digital

Approved Date: 01/2024



Purchaser's Specifications	
S.N.	
	Digital Thermometer, Clinical
	Manufacturer
	Brand
	Type/Model
	Country of Origin
1	Description of Function
1.1	Clinical digital thermometer to check the body temperature.
2	Operational Requirements
2.1	Human oral use.
3	System Configuration
3.1	Digital Thermometer, Clinical.
4	Technical Specifications
4.1	Flat type, wide thermometer, safe to use, no glass, no mercury.
4.2	Scale: Celsius scale.
4.3	Measurement range: 32°C to 45°C
4.4	Accuracy: +/- 0.1°C between 35°C to 42°C.
4.5	Display: Liquid crystal display, easy to read.
4.6	Shall work on a battery.
4.7	Shall have a low battery indicator.
4.8	Shall have facility of beep sound and switch off.
4.9	Water proof for ease of cleaning.
5	Accessories, Spares and Consumables
5.1	Shall be provided with battery. Bidder to indicate the type of and number of battery to be supplied.
5.2	Packing: Single piece packing in plastic barrel with cover.
6	Operating Environment
6.1	The system offered shall be designed to store and to operate normally under the conditions of the purchaser's country. The conditions include, Climate, Temperature, Humidity, etc.
7	Standards and Safety Requirements
7.1	CE (93/42 EEC Directives)
8	User Training
8.1	Not applicable.
9	Warranty
9.1	Warranty for 1 year after acceptance.
10	Maintenance Service During Warranty Period
10.1	Not applicable.
11	Installation and Commissioning
11.1	Must supply preassembled unit, ready to use.
12	Documentation
12.1	User's manual to be supplied in English.

Brick

[Signature]